

## DECLARATION OF CONFORMITY

### *Hazardous Area Installations (Equipment Group 2, Equipment Category 3)*

Application of Council Directives: 94/9/EC (ATEX)

Standards to which conformity declared: BS EN 60079-0:2006,  
BS EN 60079-15:2005

Manufacturer's Name and Address: Alicat Scientific, Inc.  
7641 N. Business Park Dr.  
Tucson, AZ 85743  
United States of America

Tel: 520.290.6060  
Fax: 520.290.0109  
E-Mail: [info@alicat.com](mailto:info@alicat.com)

Type of Equipment: M, MS, MC, MCS, MCP, MCR, MCRS,  
MCV, MCVS, P, PS, PC, PCS, PCD,  
PCDS, PCR, PCRS mass flow and  
pressure equipment.

Classification: Ex nA IIC, T4  
Tamb. -40°C ≤ Ta ≤ +50°C

Notifying Body: CSA International

I, the undersigned, hereby declare that the equipment listed hereinbefore conforms to the referenced directive and meets the applicable standard(s) requirement(s) contained therein



---

Jay Burhans  
Quality Manager  
Alicat Scientific, Inc.

December 2, 2013  
Date